

Teclu burner

Item no. 1300

DIN-DVGW reg. no. NG-2211AN0133

- air regulation
- needle valve
- tube Ø 19 mm
- head Ø 30 mm



Variants

Article no.	Gas
1300	Natural gas
1301	Bottle gas

DIN / Approval

Article no.	Gas
1300/1	Natural gas

Notes

Burners according to [DIN 30665](#) cannot be switched off completely. A small flame remains burning even if needle valve or stop cock are turned off. The gas flow shall be disconnected by means of the central stop cock in the laboratory.

Learn more about our laboratory burners [here](#).

The picture shows the burner with safety base (extra charge) .

Technical data

Item no.	Max. Temp. (°C)	Consumption (l/h)	Height (mm)	Weight (g)
1300	1520	200	165	330
1300/1	1550	220	165	330
1301	1600	85	165	330

Item no.	Pipe ø	Head ø	Bel. (kg)	DIN	GS	Natural gas	Propane	Safety base
1300	19	32	2,30	-	-	+	-	o
1300/1	19	32	2,30	+	+	+	-	o
1301	19	32	2,50	-	-	-	+	o