Bunsen burner stop-cock



Item no. 1030

DIN-DVGW reg. no. NG-2211AN0133

- air regulation
- stop-cock
- pilot flame inside
- adjustable for 1030

Variants

Article no.	Gas
1030	Natural gas
1031	Bottle gas

DIN / Approval

Article no.	Gas
1030/1	Natural gas
1031/1	Bottle gas

Notes

Burners according to DIN 30665 cannot be switched off completely. A small flame remains burning even if needle valve or stop cock are turned off. The gas flow shall be disconnected by means of the central stop cock in the laboratory.

Learn more about our laboratory burners here.

The picture shows the burner with safety base (extra charge).

Technical data

Item no.	Max. Temp. (°C)	Consumption (l/h)	Height (mm)	Weight (g)
1030	1500	110	155	225
1030/1	1500	110	155	225
1031	1520	50	155	225
1031/1	1520	50	155	225

Item no.	Pipe ø	Head Ø	Bel. (kg)	DIN	GS	Natural gas	Propane	Safety base
1030	14	17	1,10	–	–	+	–	o
1030/1	14	17	1,10	+	+	+	–	o
1031	14	17	1,25	–	–	–	+	o
1031/1	14	17	1,25	+	+	–	+	o