# Bunsen burner

Item no. 1000

DIN-DVGW reg. no. NG-2211AN0133

- air regulation



## DIN / Approval

| Article no. | Gas |
|---|---|
| **1000** | Natural gas |
| **1001** | Bottle gas |

## Notes

Burners according to DIN 30665 cannot be switched off completely. A small flame remains burning even if needle valve or stop cock are turned off. The gas flow shall be disconnected by means of the central stop cock in the laboratory.

Learn more about our laboratory burners here.

The picture shows the burner with safety base (extra charge).

## Technical data

| Item no. | Max. Temp. (°C) | Consumption (l/h) | Height (mm) | Weight (g) |
|---|---|---|---|---|
| **1000** | 1500 | 110 | 155 | 210 |
| **1001** | 1540 | 55 | 155 | 210 |

| Item no. | Pipe ø | Head Ø | Bel. (kg) | DIN | GS | Natural gas | Propane | Safety base |
|---|---|---|---|---|---|---|---|---|
| **1000** | 14 | 17 | 1,25 | + | + | + | – | o |
| **1001** | 14 | 17 | 1,35 | + | – | – | + | o |